

Protocol: STN POST OP 8PARRENT 9

PATIENT POSITION		IMAGING PARAMETERS	
Patient Entry	Head First	Gradient Mode	WHOLE
Patient Position	Supine	Imaging Mode	2D
Coil Configuration	HEAD	Pulse Sequence	Gradient Echo
Plane	SAGITTAL	Imaging Options	None
Series Description	SAG LOC	Acceleration Factor	1.00
SCAN TIMING		SCAN RANGE	
Flip Angle	20	FOV	30.0
TE	20.0	Slice Thickness	5.0
Number of Echoes	1	Slice Spacing	1.0
TR	250.0		
Receiver Bandwidth	19.23		
IMAGE ENHANCE		ACQ TIMING	
Filter Choice	None	Freq	256
		Phase	128
		Freq DIR	S/I
		NEX	1.00
		Phase FOV	1.00
		Auto Shim	Auto
		Phase Correction	No
GATING/TRIGGER		FMRI	
Auto Trigger Type	Off	PSD Trigger	Internal
		Slice Order	Interleaved
		View Order	Bottom/Up
		# of Repetitions REST	0
		# of Repetitions ACTIVE	0
MULTI-PHASE		SAT	
# of Acquisition	0	Tag Type	None
Seperate Series	0		
Mask Phase	0		
Mask Pause	0		
ASSET		TRICKS	
Slice Acceleration Factor	1.00	Pause On/Off	On
Phase Acceleration Factor	1.00	Auto Subtract	0
		Auto SCIC	Off
CONTRAST		MULTI-STATION	
Contrast Yes/No	No	Number of Stations	0
		Station Number	0

SAG LOC

SAG LOC

Protocol: STN POST OP 8PARRENT 9

PATIENT POSITION		IMAGING PARAMETERS	
Patient Entry	Head First	Gradient Mode	WHOLE
Patient Position	Supine	Imaging Mode	2D
Coil Configuration	HEAD	Pulse Sequence	Spin Echo
Plane	AXIAL	Imaging Options	EPI, DIFF
Series Description	DIFFUSION	PSD Name	epi2
SCAN TIMING		SCAN RANGE	
TE	Minimum	FOV	24.0
TR	10000.0	Slice Thickness	5.0
Number of Shots	1	Slice Spacing	0.0
IMAGE ENHANCE		ACQ TIMING	
Filter Choice	None	Freq	128
		Phase	192
		Freq DIR	R/L
		NEX	1.00
		Phase FOV	1.00
		Auto Shim	Auto
		Phase Correction	Yes
GATING/TRIGGER		USER CVS	
Auto Trigger Type	Off	User CV0	1.00
MULTI-PHASE		SAT	
Seperate Series	0	Tag Type	None
Mask Phase	0		
Mask Pause	0		
DIFFUSION		TRICKS	
B Value	1000	Pause On/Off	On
Optimized TE	Yes	Auto Subtract	0
Diffusion Directions	All	Auto SCIC	Off
Number of Diffusion Directions	6		
Number of T2 Images	1		
Dual Spin Echo	Off		
Recon All Images	On		
CONTRAST		MULTI-STATION	
Contrast Yes/No	No	Number of Stations	0
		Station Number	0

DIFFUSION

DIFFUSION

Protocol: STN POST OP 8PARRENT 9

PATIENT POSITION		IMAGING PARAMETERS	
Patient Entry	Head First	Gradient Mode	WHOLE
Patient Position	Supine	Imaging Mode	3D
Coil Configuration	HEAD	Pulse Sequence	SPGR
Plane	AXIAL	Imaging Options	EDR, Fast, IrP, ZIP2
Series Description	AX 3D SPGR	Acceleration Factor	1.00
SCAN TIMING		SCAN RANGE	
Flip Angle	15	FOV	30.0
TE	Min Full	Slice Thickness	2.0
TI	500	Location per Slab	128
Receiver Bandwidth	19.23	Overlap Locations	0
IMAGE ENHANCE		ACQ TIMING	
Filter Choice	None	Freq	256
		Phase	256
		Freq DIR	A/P
		NEX	1.00
		Phase FOV	1.00
		Auto Shim	Auto
		Phase Correction	No
GATING/TRIGGER		USER CVS	
Auto Trigger Type	Off	User CV23	100.00
FMRI		MULTI-PHASE	
PSD Trigger	Internal	# of Acquisition	0
Slice Order	Interleaved	Seperate Series	0
View Order	Bottom/Up	Mask Phase	0
# of Repetitions REST	0	Mask Pause	0
# of Repetitions ACTIVE	0		
SAT		ASSET	
Tag Type	None	Slice Acceleration Factor	1.00
		Phase Acceleration Factor	1.00
TRICKS		CONTRAST	
Pause On/Off	On	Contrast Yes/No	No
Auto Subtract	0		
Auto SCIC	Off		
MULTI-STATION			
Number of Stations	0		
Station Number	0		

AX 3D SPGR

AX 3D SPGR

Protocol: STN POST OP 8PARRENT 9

PATIENT POSITION		IMAGING PARAMETERS	
Patient Entry	Head First	Gradient Mode	ZOOM
Patient Position	Supine	Imaging Mode	2D
Coil Configuration	HEAD	Pulse Sequence	Gradient Echo
Plane	AXIAL	Imaging Options	FC
Series Description	AX MPGR	Acceleration Factor	1.00
SCAN TIMING		SCAN RANGE	
Flip Angle	20	FOV	22.0
TE	20.0	Slice Thickness	5.0
Number of Echoes	1	Slice Spacing	1.0
TR	600.0		
Receiver Bandwidth	20.83		
IMAGE ENHANCE		ACQ TIMING	
Filter Choice	None	Freq	256
		Phase	192
		Freq DIR	A/P
		NEX	2.00
		Phase FOV	0.75
		Auto Shim	Auto
		Phase Correction	No
GATING/TRIGGER		FMRI	
Auto Trigger Type	Off	PSD Trigger	Internal
		Slice Order	Interleaved
		View Order	Bottom/Up
		# of Repetitions REST	0
		# of Repetitions ACTIVE	0
MULTI-PHASE		SAT	
# of Acquisition	0	Tag Type	None
Seperate Series	0		
Mask Phase	0		
Mask Pause	0		
ASSET		TRICKS	
Slice Acceleration Factor	1.00	Pause On/Off	On
Phase Acceleration Factor	1.00	Auto Subtract	0
		Auto SCIC	Off
CONTRAST		MULTI-STATION	
Contrast Yes/No	No	Number of Stations	0
		Station Number	0

AX MPGR

AX MPGR

Protocol: STN POST OP 8PARRENT 9

PATIENT POSITION		IMAGING PARAMETERS	
Patient Entry	Head First	Gradient Mode	WHOLE
Patient Position	Supine	Imaging Mode	3D
Coil Configuration	HEAD	Pulse Sequence	SPGR
Plane	AXIAL	Imaging Options	EDR, Fast, IrP, ZIP2
Series Description	OPTION AX 3D STEALTH	Acceleration Factor	1.00
SCAN TIMING		SCAN RANGE	
Flip Angle	15	FOV	30.0
TE	Min Full	Slice Thickness	2.0
TI	500	Location per Slab	90
Receiver Bandwidth	16.67	Overlap Locations	0
IMAGE ENHANCE		ACQ TIMING	
Filter Choice	None	Freq	256
		Phase	256
		Freq DIR	A/P
		NEX	1.00
		# of Acq. Before Pause	0
		Phase FOV	1.00
		Auto Shim	Auto
		Phase Correction	No
GATING/TRIGGER		USER CVS	
Auto Trigger Type	Off	User CV23	100.00
FMRI		MULTI-PHASE	
PSD Trigger	Internal	# of Acquisition	0
Slice Order	Interleaved	Seperate Series	0
View Order	Bottom/Up	Mask Phase	0
# of Repetitions REST	0	Mask Pause	0
# of Repetitions ACTIVE	0		
SAT		ASSET	
Tag Type	None	Slice Acceleration Factor	1.00
		Phase Acceleration Factor	1.00
TRICKS		CONTRAST	
Pause On/Off	On	Contrast Yes/No	No
Auto Subtract	0		
Auto SCIC	Off		
MULTI-STATION			
Number of Stations	0		
Station Number	0		

OPTION AX 3D STEALTH

OPTION AX 3D STEALTH